



PELVI.LOC® PL-3DASR -XS, -S, -M, -L



Functioning under the cover see image below



For the safe use of this product, read and observe these assembly instructions and the manual before the first use.



Intended use

The **PELVI.LOC®** PL-3DASR is intended for people with restricted mobility, especially for pelvic positioning within the wheelchair and active-wheelchair and prevents from sliding forward.

Intended Patient Group:

- Infants, adolescents and adults
- Sex: female and male

Receipt Control

Completely unpack the product and check the delivery for completeness and transport damages. Should you – against expectation – find a defect in the delivered product then:

1. Immediately lodge a complaint with the freight carrier
2. Have the damage/defect recorded
3. Notify us or your retailer immediately

Scope of Delivery

- 1) 1 x **PELVI.LOC**® PL-3DASR -XS, -S, -M, -L, not pre-assembled
- 2) 1 x Information for online download of the manual

Combination of medical devices

Suitable for medical devices which are intended for the use of a pelvis positioning belt

Assembly



Assembly should be carried out by a person with specialist knowledge.

1. Place the paddings (a) on the pelvic area / antecubital fold of the hip with the ratchet (b) facing upwards.
 - Afterwards, the logo on the padding should be readable from the front side!
2. Set the spacing of the paddings using the hole strap (d) – The pelvis-retraction-belt is tightened by operating the buckle (e)
 - The hinge of the ratchet should be centred between the paddings.
 - Pull out the ratchet not further than to the point, where two teeth are still free behind each catch of the buckle (e).
 - The Velcro can be released for adjustments and the padding can be opened
3. After that, attach the ratchet (b) to the frame tube of the wheelchair. (observe installation notes)
4. Tighten the ratchet (b) on both sides equally to a degree that still allows enough space for a flat hand between the thigh and the paddings.
5. Both ends of the ratchet can be cut to length in a semicircle, if required. Deburr the ends of the cut ratchet, if necessary. –Danger of injury!
6. Check the firm fit and locking on both sides of the ratchet!



Manual



Safety Notes

In spite of their stable construction, components like the hole straps, webbings, ratchets and paddings are subject to wear and tear. It is therefore recommended to perform a thorough inspection of the **PELVI.LOC**® every 3 Months. It is thereby important to make sure the rivet and screw connections (d) are firmly secure!

- Before using the product, ensure that it is in a serviceable condition and that it fits correctly. – Danger of accidents!
- Do not patch up and/or glue up and reuse damaged components such as hole stays (f), ratchets (b) or paddings (a). – Danger of accidents!
- Do not sit on the paddings – danger of bruising, especially by the catches of the buckles (c)!
- Do not lift the wheelchair by the **PELVI.LOC®**. Danger of accidents!
- Be careful that your hands do not get caught between the webbing and the frame or seat tube. Danger of crushing!

Restrictions on use / undesirable side effects

- Do not use the **PELVI.LOC®** if there are any bruises, wounds or other not well-healed injuries in the area where the paddings will be. – Danger of injury!
- Do not use the **PELVI.LOC®** in case of osteoporosis, brittle bone disease or severe circulatory disorders. – Danger of injury!
- After cleaning, ensure the correct functioning of the **PELVI.LOC®**!
- A damaged **PELVI.LOC®** must be replaced immediately. – Danger of accidents!



Attention:

- The **PELVI.LOC®** is not part of the restraint system for the wheelchair and/or user during transport in a vehicle for the transport of disabled persons.
- Neither the manufacturer nor the distributor will be liable, in the sense of product liability, for damages or risks caused by unattended children with regard to the product!

Handling

1. To use the **PELVI.LOC®** place the paddings (a) on the pelvic area / antecubital fold of the hip



Caution: Make sure that there are no items under the padding! This will prevent painful bruises.

2. After that, guide the ratchet (b) through the corresponding catch of the buckle. For that, operate the corresponding pressure buckle (e).
 3. Then, tighten the ratchet (b) to a degree that still allows enough space for a flat hand between the pelvic area and the padding.
 4. Make sure that the catches of the buckles (c) engage with the teeth of the ratchet!
- Check the tension!

In case of an accidental loosening the **PELVI.LOC®** positioning system can be tightened again.

To open the **PELVI.LOC®** operate the catch of the buckle (c) and pull out the ratchet (b).

- Make sure that the paddings do not get caught between the stokes of the driving wheel!

Cleaning and Disinfection

- Machine washable up to 40 degrees.
- Thermochemical disinfection possible by adding ELTRA 40 when washing
- No spinning
- Not dryer-proof

Maintenance

Inspect the **PELVI.LOC**[®] for damages and check the functioning on a regular basis.

Repair

Immediately replace defective **PELVI.LOC**[®] systems.

Disposal

- The packaging is recyclable.
- The metal parts can be disposed of through scrap metal recycling.
- Plastic parts are suitable for plastic recycling.
- The disposal must be made according to national statutory provisions.
- Please inquire with your city or municipal administration about local disposal services.



Technical Data

	XS	S	M	L
Seat width (cm)	16 - 24	20 - 32	30 - 40	36 - 46
Padding size (cm)	17,5 x 9	20 x 11	23 x 13	25,5 x 14
Weight (g)	220	220	280	280
Predetermined breaking point (at approx. kg)	60	60	80	80

Ratchet size XS - S	1,9 x 0,2 x 40,0 cm
Ratchet size M - L	2,4 x 0,2 x 40,0 cm

Product is
latex-free



Warranty

We provide warranty for this product in accordance with the legal provisions. This warranty includes material and processing defects. This does not include wear parts and components/assemblies that are subject to intended wear and tear as well as damages that are the result of excessive strain, unintended use, forcible damage or unauthorised modification/repair. If you wish to make a warranty claim, please inquire with the retailer you purchased the product from. For product liability reasons, repairs and maintenance work must only be conducted by authorised speciality retailers, using original spare parts.



Manufacturer:

A2J GmbH

Am Fuchsberg 13

D-87452 Altusried

Fon: +49 (08373) 987269

Fax: +49 (08373) 987276

eMail: info@a2j-intl.com

Internet: www.a2j-intl.com

As of 01.06.2019



This product complies with the essential requirements of the EC Directive 93/42/EEC concerning medical devices

Subject to technical modifications